

MEASURING AND CONTROL EQUIPMENT ON THE RFSOC FOR HOBICAT FACILITY

Wednesday, October 25, 2023 2:34 PM (4 minutes)

Controlling SRF cavities in CW mode in the presence of mechanical disturbances, as well as in the presence of noisy detectors, makes control still a challenge. Internal cavity features such as Lorentzian forces are well understood and predictable, but also make control difficult. An inexpensive solution for compact accelerators that can accommodate many algorithms simultaneously on a single board at an affordable price is RFSoc. Their analog circuits are not yet accurate. Nevertheless, a significant reduction in development time can be achieved through an integrated architecture, as well as development tools available on the market.

The following will present the RFSoc control hardware with a focus on basic functions, implemented primarily as a firmware solution. The scope of functions is as follows: VNA measurements to determine resonator quality factor and S-parameters, PLL, control in the presence of noisy detector data, and adaptive RF/piezo control. The results of their tests on the Hobicat test facility will be shown.

Keyword

RFSoc

Primary authors: USHAKOV, Andriy (Helmholtz-Zentrum Berlin); Dr NEUMANN, Axel (Helmholtz Zentrum Berlin); Dr ECHEVARRIA FERNANDEZ, Pablo (Helmholtz Zentrum Berlin)

Session Classification: Posters

Track Classification: Measurement and control