

Design of an X-ray Ionization Beam Profile Monitor for Korea-4GSR

Thursday, November 14, 2024 2:55 PM (15 minutes)

A photon beam generated by the Insertion Device (ID) of a synchrotron light source can be contaminated by radiation from upstream and downstream bending magnets, leading to position measurement errors in blade-type monitors. The operation of the Korea-4GSR, which has extremely low emittance, is particularly sensitive to photon beam position variations, necessitating more accurate position measurements. To robustly measure the position and simultaneously obtain the profile of a photon beam in a non-destructive manner, we are developing an ionization profile monitor.

We designed a noble gas environment to ensure adequate signal strength and incorporated a defocusing electrode structure to fully utilize the relatively large active area of the readout. Since magnification in the defocusing field depends on the vertical position, we proposed a calibration method to correct the non-linearity, which we then verified through particle tracking simulation.

Best Presentation

Yes

Paper submission Plan

No

Contribution track

ICABU WG2. Beam Physics, Diagnostics & Novel Techniques

Primary author: SONG, Woojin (POSTECH)

Co-authors: HAHN, Garam (Pohang Accelerator Laboratory, POSTECH); HYUN, Hyojung (Pohang Accelerator Laboratory); Dr JEHAN, Kim (Pohang Accelerator Laboratory); Dr JAE-HONG, Lim (Pohang Accelerator Laboratory); Prof. CHUNG, Moses (POSTECH); Dr YOUNGSUNG, Park (Pohang Accelerator Laboratory); MIN-HO, Seo (Pohang Accelerator Laboratory); KIM, Seonghan (Pohang Accelerator Laboratory); LEE, Seungcheol (Pohang Accelerator Laboratory); HWANG, Sunmin

Presenter: SONG, Woojin (POSTECH)

Session Classification: ICABU WG2

Track Classification: ICABU: ICABU WG2. Beam Physics, Diagnostics & Novel Techniques