The 26th International Conference on Accelerators and Beam Utilizations

Contribution ID: 76

Type: Poster

Design of a compact gantry with high energy acceptance for carbon-ion cancer therapy

Thursday, November 14, 2024 1:00 PM (1h 30m)

We presented gantry design with compact size and high energy acceptance satisfying rotation invariant condition. three identical 90 degree superconducting canted cosine theta (CCT) bending magnets (BMs) was used in a gantry to make gantry size be small, however, one BM have reverse bending angle. For high energy acceptance, low dispersion should be maintained through a gantry, therefore, alternating-grediant quadrupole coils was introduced in each 90 degree BM. Because rotator beam line was not intorduced in this gantry, pointto-parallel imaging in the entire gantry is satisfied to ensure rotation invariant optics. Downstream scanning method is adapted in this gantry in order to ensure small nonlinear beam dynamics in a last BM by making the BM's bore radius be small. 90 degree CCT BM design was not fullfilled yet and it wii be future work.

Paper submission Plan

Yes

Best Presentation

No

Contribution track

ICABU WG1. Accelerator Systems

Primary author: OH, bonghoon (korea university)
Co-author: KO, Jinjoo (Korea university)
Presenter: OH, bonghoon (korea university)
Session Classification: ICABU Poster Session

Track Classification: ICABU: ICABU WG1. Accelerator Systems